


# Schedule of Accreditation

issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>0232</b>	<b>Indentec Hardness Testing Machines Ltd</b>	
	<b>Issue No: 13</b>	<b>Issue date: 14 February 2003</b>
	<b>Lye Valley Industrial Estate Bromley Street Lye Stourbridge West Midlands DY9 8HX</b>	<b>Contact: Mr J Piller Tel: +44 (0) 1384 896949 Fax: +44 (0) 1384 424470 E-Mail: <a href="mailto:mail@indentec.demon.co.uk">mail@indentec.demon.co.uk</a> Website: <a href="http://www.indentec.demon.co.uk">http://www.indentec.demon.co.uk</a></b>

## SUMMARY OF ACCREDITATION

Calibration performed on permanent laboratory premises

### HARDNESS

Indentors, ball  
Test blocks, Rockwell  
Test blocks, Vickers  
Test blocks, Brinell  
Testing machines, Rockwell, direct  
Testing machines, Rockwell, indirect  
Testing machines, Vickers, direct  
Testing machines, Vickers, indirect  
Testing machines, Brinell, direct  
Testing machines, Brinell, indirect



Calibration performed on permanent laboratory premises

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

### Indentec Hardness Testing Machines Ltd

Issue No: 13 Issue date: 14 February 2003

#### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
<b>HARDNESS</b>			
Calibration of Rockwell Standardized Hardness Blocks	Rockwell scales: A, B, C, D, E, F, G, H, K, N and T	1 Rockwell Unit 1	<p>1 The calibration/verification shall be in accordance with the requirements of BS EN ISO 6508:Parts 1, 2 and 3:1999 and ASTM E18-02</p> <p>2 The verification shall be in accordance with the requirements of BS EN ISO 6507:1998 and ASTM E92-97.</p> <p>3 The verification shall be in accordance with the requirements of BS EN ISO 6506:1998 and ASTM E10-01.</p>
Calibration of Vickers Reference Hardness Blocks	Vickers scales: HV 50 to HV 1	1% HV 2	
Calibration of Brinell Reference Hardness Blocks	Brinell scales: From HBW 10/3000 to HBW 1/1	0.5% on indentation size mm 3	
Certification of Ball Indenters	Rockwell	See Note 1	
	Brinell	See Note 3	
<b>CERTIFICATION OF HARDNESS MEASURING MACHINES IN SERVICE</b>			
Direct verification of Rockwell Hardness Testing Machines	Rockwell scales: A, B, C, D, E, F, G, H, K, N and T	0.12% force 0.1 $\mu$ m length 0.1 second time	
Indirect verification of Rockwell Hardness Testing Machines	Rockwell scales: A, B, C, D, E, F, G, H, K, N and T	1 Rockwell unit 1	
Direct verification of Vickers hardness calibrating and testing machines	Vickers scales: HV 50 to HV 1	0.12% force 0.1 second time	
Verification of indentation measuring equipment for Vickers hardness	1.0 mm to 0.51 mm 0.5 mm to 0.05 mm 1.0 mm to 0.05 mm	Direct 2 $\mu$ m Direct 1 $\mu$ m Indirect 0.5% (in mm) on diagonal (2 m minimum)	
Indirect verification of Vickers hardness testing machines	Vickers scales: HV 5 to HV 100 HV 0.2 to HV 3	1% HV	
Direct verification of Brinell Hardness Testing Machines	Brinell scales: From HBW 10/3000 to HBW 1/1	0.24% force 0.1 second time	
Indirect verification of Brinell Hardness Testing Machines	Brinell scales: From HBW 10/3000 to HBW 1/1	1.0% HBW	
Verification of indentation measuring equipment for Brinell hardness	Direct	0.1% or 10 $\mu$ m whichever is the larger	
END			